



CERTIFICATE OF ANALYSIS

Histamine Check Swab

Product code :60448

Lot No.: 20240325J

Expiry date: Sep. 24, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Absorbance(600nm) of 100ppm histamine solution	1.67	1.55-1.95

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Ryosuke Sugimoto; Manager

Date Jun. 24, 2024

KBG6044800aa-Eaa
SA2406-0219

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company