

**CERTIFICATE OF ANALYSIS****CheckLite 250 Plus**

Product code :61305

Lot No.: 20240610A

Expiry date: Oct. 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Luciferin-luciferase reagent Luminescence:Std.(%)	101.5	100±15
Reconstitution buffer for luciferin-luciferase reagent Background ATP( $\times 10^{-13}$ M)	0.5	$\leq 4.0$
ATP releasing reagent Background ATP( $\times 10^{-13}$ M)	0.1	$\leq 4.0$

Storage condition : at 2-8°C, protected from light

This is to certify that above information is a true accurate copy of data obtained for this kit.

  
Ryosuke Sugimoto; ManagerDate Jun. 24, 2024KBG6031201ae-Eaa  
SA2406-0218Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company