



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20240913U

Expiry date: Dec. 12, 2025

TEST PARAMETER	RESULT	SPECIFICATION
Average Luminescence(RLU)	13400	$\geq 9100, \leq 16900$
(When applied 100μL of 5 pmol AMP: measured with Lumitester Smart)		

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama
Hiroshi Katayama; Manager

Date Oct. 10, 2024

KBG6033100ab-Eaa
SA2410-0147

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company