

QC certificate

- Certificate of Analysis -

Product name:	Procheck E-W
Lot number:	KM5I111D
Expire date:	2026/09

[1] Physical Characteristics

Appearance (cleanness)	clear
Appearance (color)	yellow

[2] QC test result

Sensitivity is good enough according to our QC standard.

Detection limit using bovine serum albumin is 0.02 mg

Turns blue or green	
0.16mg	+
0.08mg	+
0.04mg	+
0.02mg	+

Name of QC responsible person: Yukari Kataoka

Date: 12-Sep-2024

Signature of responsible person Yukari Kataoka

Nikken Bio Inc.
Teigaiberi 23 Ohashiberi Kumiyama-cho Kuze-gun KYOTO 613-0046 JAPAN

