

**CERTIFICATE OF ANALYSIS****LuciPac A3 Surface (Pre-Moistened)**

Product code :60367

Lot No.: 20240917M

Expiry date: Sep. 16, 2025

| <u>TEST PARAMETER</u>       | <u>RESULT</u> | <u>SPECIFICATION</u>     |
|-----------------------------|---------------|--------------------------|
| Average Luminescence* (RLU) | 17820         | $\geq 10500, \leq 19500$ |

\* When applied 100 $\mu$ L of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hiroshi Katayama, Manager

Date Oct. 22, 2024

KBG6036700ab-Eaa  
SA2410-0272

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company