

**CERTIFICATE OF ANALYSIS****LuciPac Pen**

Product code :60331

Lot No.: 20240927U

Expiry date: Dec. 26, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	14200	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama  
Hiroshi Katayama/Manager

Date Oct. 22, 2024KBG6033100ab-Eaa  
SA2410-0265Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company