

CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

PRODUCT	LOT NUMBER	EXPIRY DATE	CERTIFICATION DATE
MW570PL	073183	31/03/2025	16/04/2024

Listeria Isolation Transwab®

PARAMETER	SPECIFICATION	RESULT														
Sterilisation Method	Irradiation	PASSED														
Sterility Test	Visual inspection following incubation	PASSED														
Performance Testing	<p>Colour change induced by growth of following organisms:</p> <table><tr><td><i>Listeria monocytogenes</i></td><td>NCTC 5214 (24hr)</td></tr><tr><td><i>Listeria monocytogenes</i></td><td>NCTC 7993 (24hr)</td></tr><tr><td><i>Listeria monocytogenes</i></td><td>NCTC 11994 (24 hr)</td></tr><tr><td><i>Listeria innocua</i></td><td>NCTC 11288 (24 hr)</td></tr></table> <p>No colour change induced by growth of following organisms</p> <table><tr><td><i>Staphylococcus aureus</i></td><td>NCTC 8532 (24hr)</td></tr><tr><td><i>Listeria grayii</i></td><td>NCTC 10815 (24 hr)</td></tr><tr><td><i>Enterococcus faecalis</i></td><td>ATCC® 19433 (24hr)</td></tr></table>	<i>Listeria monocytogenes</i>	NCTC 5214 (24hr)	<i>Listeria monocytogenes</i>	NCTC 7993 (24hr)	<i>Listeria monocytogenes</i>	NCTC 11994 (24 hr)	<i>Listeria innocua</i>	NCTC 11288 (24 hr)	<i>Staphylococcus aureus</i>	NCTC 8532 (24hr)	<i>Listeria grayii</i>	NCTC 10815 (24 hr)	<i>Enterococcus faecalis</i>	ATCC® 19433 (24hr)	PASSED
<i>Listeria monocytogenes</i>	NCTC 5214 (24hr)															
<i>Listeria monocytogenes</i>	NCTC 7993 (24hr)															
<i>Listeria monocytogenes</i>	NCTC 11994 (24 hr)															
<i>Listeria innocua</i>	NCTC 11288 (24 hr)															
<i>Staphylococcus aureus</i>	NCTC 8532 (24hr)															
<i>Listeria grayii</i>	NCTC 10815 (24 hr)															
<i>Enterococcus faecalis</i>	ATCC® 19433 (24hr)															

Product has been manufactured in accordance with ISO 9001:2015

Certificate LRQ0939857 B

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance

