

# CERTIFICATE OF ANALYSIS

## Easy Plate AC

Product code :61973

Lot No.: AC20240726A

Expiry date: Jan. 23, 2026

| <u>TEST PARAMETER</u>      | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test           | passed test   | * See below          |
| Spreading time             | passed test   | $\leq$ 60 sec        |
| Gelling of medium          | passed test   | Gelling within 3 min |
| Curl of the base sheet     | passed test   | $\leq$ 4 mm          |
| Appearance of culture area | passed test   | Light yellow         |
| Scanning of QR code        | passed test   | Passed               |

\*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 70-200

Bacillus subtilis (NBRC3134) : 70-350

Klebsiella pneumoniae (NBRC14940) : 50-200

Pseudomonas aeruginosa (NBRC3899) : 70-350

Staphylococcus aureus (ATCC25923) : 50-200

Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8℃

H Katayama  
Hiroshi Katayama, Manager

Date Nov. 25, 2024

KBG6197300aa-Eaa  
SA2411-0272

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company