

CERTIFICATE OF ANALYSIS

Easy Plate YM-R

Product code :61977

Lot No.: YM20240827A

Expiry date: Feb. 26, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	≤ 60 sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	≤ 4 mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

Performance test

Aspergillus niger (NBRC 33023) : 50-500*
 Penicillium citrinum (NBRC 6352) : 50-500*
 Candida albicans (NBRC 1594) : 50-500*
 Saccharomyces cerevisiae (NBRC 100929) : 50-500*
 Bacillus subtilis (NBRC 3134) : No growth
 Escherichia coli (NBRC 15034) : No growth
 Pseudomonas aeruginosa (NBRC 3899) : No growth

*The ratio of the average colony count on the product to the potato dextrose agar (PDA) medium(%).

Storage condition : 2-8°C


 Hiroshi Katayama; Manager

Date Nov. 25, 2024

KBG6197700aa-Eaa
 SA2411-0280

Quality Assurance Group, Edogawa
 Quality Unit
 Kikkoman Biochemifa Company