

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20241211Y

Expiry date: Mar. 10, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13300	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama
Hiroshi Katayama, Manager

Date Dec. 19, 2024

KBG6036100ac-Eaa
SA2412-0193

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company