

CERTIFICATE OF ANALYSIS

Histamine Test

Product code :61341

Lot No.: 20240912A

Expiry date: Apr. 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	100.3	100±5
Absorbance of reagent blank	0.011	≤0.04
Histamine dehydrogenase activity(U/vial)	2.22	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	97.8	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama
Hiroshi Katayama, Manager

Date Dec. 19, 2024

KBG6044101aa-Eaa
SA2412-0197

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company