

CERTIFICATE OF ANALYSIS**LuciPac A3 Water**

Product code :60365

Lot No.: 20241126W

Expiry date: Feb. 25, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12200	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H. Katayama
Hiroshi Katayama, Manager

Date Dec. 19, 2024

KBG6036500ac-Eaa
SA2412-0195

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company