



Kikkoman Biochemifa Company

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20250108Y

Expiry date: Apr. 07, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	13000	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13200	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Id. Kikuchi
Hisako Kikuchi; Manager

Date Jan. 21, 2025

KBG6036100ac-Eaa
SA2501-0152

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company