



Kikkoman Biochemifa Company

**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Water**

Product code :60365

Lot No.: 20250106W

Expiry date: Apr. 05, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13400	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi  
Hisako Kikuchi; Manager

Date Jan. 21, 2025

KBG6036500ac-Eaa  
SA2501-0153

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company