

## CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

PRODUCT	LOT NUMBER	EXPIRY DATE	CERTIFICATION DATE
MW572M 125	083129	30/06/2025	26/06/2024

### Salmonella Isolation Transwab®

PARAMETER	SPECIFICATION	RESULT
Sterilisation Method	Irradiation	PASSED
Sterility Test	Visual inspection following incubation	PASSED
Performance Testing	<p>Colour change induced by growth of following organisms:</p> <p><i>Salmonella enteritidis subsp enterica</i> NCTC 4444</p> <p><i>Salmonella typhimurium</i> NCTC 0074</p> <p>No colour change induced by growth of following organisms</p> <p><i>Escherichia coli</i> NCTC 9001</p> <p><i>Staphylococcus aureus</i> NCTC 8532 (24hr)</p> <p><i>Enterococcus faecalis</i> ATCC® 19433 (24hr)</p>	PASSED

Product has been manufactured in accordance with ISO 9001:2015

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance