

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361


Lot No.: 20250203Y

Expiry date: May. 02, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Feb. 17, 2025

KBG6036100ac-Eaa
SA2502-0169

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company