

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20250307Y

Expiry date: Jun. 06, 2026

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|--|---------------|-------------------------|
| When applied 100μL of 5pmol ADP Average Luminescence(RLU) | 12600 | $\geq 9100, \leq 16900$ |
| When applied 100μL of 5pmol AMP Average Luminescence(RLU) | 12700 | $\geq 9100, \leq 16900$ |

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi
Hisako Kikuchi; Manager

Date Mar. 18, 2025

KBG6036100ac-Eaa
SA2503-0227

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company