

CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365


Lot No.: 20250127W

Expiry date: Apr. 26, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11200	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	11900	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date Mar. 18, 2025

KBG6036500ac-Eaa
SA2503-0228

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company