

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan
TEL: 81-3-5229-7911 FAX: 81-3-5261-4130
URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Swab Test	Code No.	ST-25PBS
Volume	10ml	Lot No.	4A4600K
Manufactured	Jun 21, 2024	Doc. No.	CAST25PBS-240805-2
Expiration	May. 2026	Doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Appearance: Swab	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	10.2g±0.2g	10.2 g	passed
NaCl Content	0.85±0.1%	0.81%	passed
pH	6.6~7.6	7.2	passed
Leakage(full inspection)	no leak, -0.04Mpa hold 15seconds.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024