

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan

TEL: 81-3-5229-7911 FAX: 81-3-5261-4130

URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G1884
Manufactured	Jun 03, 2024	doc. No.	CAMT11PBS-20240805-1
Expiration	May. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	9.1 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan

TEL: 81-3-5229-7911 FAX: 81-3-5261-4130

URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G1984
Manufactured	Jun 11, 2024	doc. No.	CAMT11PBS-20240805-2
Expiration	May. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.9 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan

TEL: 81-3-5229-7911 FAX: 81-3-5261-4130

URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G2084
Manufactured	Jun 18, 2024	doc. No.	CAMT11PBS-20240805-3
Expiration	May. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	9.0 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.3	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G2184
Manufactured	Jul 03, 2024	doc. No.	CAMT11PBS-20240805-4
Expiration	Jun. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.8 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024
	

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan

TEL: 81-3-5229-7911 FAX: 81-3-5261-4130

URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G2284
Manufactured	Jul 05, 2024	doc. No.	CAMT11PBS-20240805-5
Expiration	Jun. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	9.0 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024

ELMEX LTD

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan

TEL: 81-3-5229-7911 FAX: 81-3-5261-4130

URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	4G2384
Manufactured	Jul 08, 2024	doc. No.	CAMT11PBS-20240805-6
Expiration	Jun. 2026	doc. Issued	Aug 05, 2024

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.8 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.2	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
August 5, 2024	August 5, 2024