

CERTIFICATE OF ANALYSIS

LuciPac A3 Surface

Product code :60361

Lot No.: 20250403Y

Expiry date: Jul. 02, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	13000	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

J. d. Kikuchi

Hisako Kikuchi; Manager

Date Apr. 18, 2025

KBG6036100ac-Eaa
SA2504-0224

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company