

# CERTIFICATE OF ANALYSIS

## LuciPac A3 Water

Product code :60365

Lot No.: 20250331W

Expiry date: Jun. 30, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12600	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12400	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

Hisako Kikuchi  
Hisako Kikuchi; Manager

Date Apr. 18, 2025

KBG6036500ac-Eaa  
SA2504-0226

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company