

CERTIFICATE OF ANALYSIS

Histamine Test

Product code :61341

Lot No.: 20241118A

Expiry date: Jun. 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	100.6	100±5
Absorbance of reagent blank	0.012	≤0.04
Histamine dehydrogenase activity(U/vial)	2.23	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	99.8	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Apr. 18, 2025

KBG6044101aa-Eaa
SA2504-0229

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company