

CERTIFICATE OF ANALYSIS

Easy Plate AC

Product code :61973

Lot No.: AC20241030A

Expiry date: Apr. 29, 2026

| <u>TEST PARAMETER</u> | <u>RESULT</u> | <u>SPECIFICATION</u> |
|----------------------------|---------------|----------------------|
| Performance test | passed test | * See below |
| Spreading time | passed test | ≤ 60 sec |
| Gelling of medium | passed test | Gelling within 3 min |
| Curl of the base sheet | passed test | ≤ 4 mm |
| Appearance of culture area | passed test | Light yellow |
| Scanning of QR code | passed test | Passed |

*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 70-200

Bacillus subtilis (NBRC3134) : 70-350

Klebsiella pneumoniae (NBRC14940) : 50-200

Pseudomonas aeruginosa (NBRC3899) : 70-350

Staphylococcus aureus (ATCC25923) : 50-200

Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8°C

Hisako Kikuchi
Hisako Kikuchi; Manager

Date Apr. 18, 2025

KBG6197300aa-Eaa
SA2504-0231

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company