

# CERTIFICATE OF ANALYSIS

## Easy Plate AC

Product code :61973

Lot No.: AC20241121A

Expiry date: May. 20, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	* See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

\*The ratio of the average colony count on the product to the plate count agar(%).

Escherichia coli (NBRC15034) : 70-200

Bacillus subtilis (NBRC3134) : 70-350

Klebsiella pneumoniae (NBRC14940) : 50-200

Pseudomonas aeruginosa (NBRC3899) : 70-350

Staphylococcus aureus (ATCC25923) : 50-200

Enterococcus faecalis (NBRC100482) : 70-200

Storage condition : 2-8°C



Hisako Kikuchi; Manager

Date Apr. 18, 2025

KBG6197300aa-Eaa  
SA2504-0233

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company