

**ELMEX LTD**

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan  
TEL: 81-3-5229-7911 FAX: 81-3-5261-4130  
URL <http://www.elmex.co.jp>

## Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	5G0784
Manufactured	Feb 27, 2025	doc. No.	CAMT11PBS-20250408
Expiration	Jan. 2027	doc. Issued	Apr 08, 2025

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	8.8 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.1	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
April 8, 2025	April 8, 2025