

## CERTIFICATE OF ANALYSIS

### Easy Plate SA

Product code :61976

Lot No.: SA20250127A

Expiry date: Jan. 26, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Performance test	passed test	See below
Spreading time	passed test	$\leq 60$ sec
Gelling of medium	passed test	Gelling within 3 min
Curl of the base sheet	passed test	$\leq 4$ mm
Appearance of culture area	passed test	Light yellow
Scanning of QR code	passed test	Passed

#### Performance test

Staphylococcus aureus (ATCC25923) : 50-200 (as blue colony)<sup>\*1</sup>

Staphylococcus aureus (NBRC13276) : 50-200 (as blue colony)<sup>\*1</sup>

Enterococcus faecalis (NBRC100482) : No growth

Bacillus subtilis (NBRC3134) : No growth

Escherichia coli (NBRC15034) : No growth


Proteus mirabilis (NBRC105697) : No growth

Bacillus licheniformis (NBRC12200) : Growth as pink colony

Staphylococcus epidermidis (NBRC100911) :  $\leq 10$  and no growth as blue colony<sup>\*2</sup>
<sup>\*1</sup>The ratio of the average colony count on the product to the plate count agar (%).

<sup>\*2</sup>The ratio of the average colony count on the product to Easy Plate AC (%).

Storage condition : 2-8°C

  
 Hisako Kikuchi; Manager

Date May. 21, 2025

KBG6197600aa-Eab  
SA2505-0137

Quality Assurance Group, Edogawa  
 Quality Unit  
 Kikkoman Biochemifa Company