

# CERTIFICATE OF ANALYSIS

## LuciPac A3 Surface

Product code :60361


Lot No.: 20250425Y

Expiry date: Jul. 24, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	13200	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	13400	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date May. 21, 2025

KBG6036100ac-Eaa  
SA2505-0135

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company