



Kikkoman Biochemifa Company

**CERTIFICATE OF ANALYSIS**

**LuciPac A3 Surface**

Product code :60361


Lot No.: 20250422Y

Expiry date: Jul. 21, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12100	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date May. 21, 2025

KBG6036100ac-Eaa  
SA2505-0142

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company