

## CERTIFICATE OF STERILITY AND QUALITY ASSURANCE

| PRODUCT    | LOT NUMBER | EXPIRY DATE | CERTIFICATION DATE |
|------------|------------|-------------|--------------------|
| MW572M 125 | 122184     | 31/03/2026  | 03/04/2025         |

### Salmonella Isolation Transwab®

| PARAMETER            | SPECIFICATION   | RESULT |
|----------------------|---|--------|
| Sterilisation Method | Irradiation   | PASSED |
| Sterility Test       | Visual inspection following incubation  | PASSED |
| Performance Testing  | <p>Colour change induced by growth of following organisms:</p> <p><i>Salmonella enteritidis subsp enterica</i> NCTC 4444</p> <p><i>Salmonella typhimurium</i> NCTC 0074</p> <p>No colour change induced by growth of following organisms</p> <p><i>Escherichia coli</i> NCTC 9001</p> <p><i>Staphylococcus aureus</i> NCTC 8532 (24hr)</p> <p><i>Enterococcus faecalis</i> ATCC® 19433 (24hr)</p> | PASSED |

Product has been manufactured in accordance with ISO 9001:2015

Performance testing conducted in accordance with: Internal Procedures

For and on behalf of Medical Wire & Equipment



Quality Assurance