

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface**

Product code :60361


Lot No.: 20250521Y

Expiry date: Aug. 20, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13500	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Jun. 05, 2025KBG6036100ac-Eaa
SA2506-0039Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company