

Certificate of Analysis

Product	Swab Test	Code No.	ST-25PBS
Volume	10ml	Lot No.	5A3500K
Manufactured	May 23, 2025	Doc. No.	CAST25PBS-250630
Expiration	Apr. 2027	Doc. Issued	Jun 30, 2025

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Appearance: Swab	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	10.2g±0.2g	10.2 g	passed
NaCl Content	0.85±0.1%	0.83%	passed
pH	6.6~7.6	7.1	passed
Leakage(full inspection)	no leak, -0.04Mpa hold 15seconds.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
June 30, 2025	June 30, 2025
	