

# CERTIFICATE OF ANALYSIS

## LuciPac A3 Water

Product code :60365

Lot No.: 20250602W

Expiry date: Sep. 01, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	11900	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12000	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Jun. 19, 2025

KBG6036500ac-Eaa  
SA2506-0205

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company