

CERTIFICATE OF ANALYSIS**LuciPac A3 Surface (Pre-Moistened)**

Product code :60367

Lot No.: 20250526N2

Expiry date: May. 25, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence* (RLU)	18230	$\geq 10500, \leq 19500$

* When applied 100 μ L of AMP 2.5 pmol and ADP 2.5 pmol (total 5 pmol): measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; ManagerDate Jun. 19, 2025KBG6036700ab-Eaa
SA2506-0206Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company