

CERTIFICATE OF ANALYSIS

LuciSwab ES 3.7-4.2

Product code :60357

Lot No.: 25YE1212C02

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Blank Luminance Assay (RLU)		
Average Luminescence	passed test	≤ 20
Maximum Luminescence	passed test	≤ 25

Measured with LuciPac A3 Water and Lumitester Smart

This is to that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Jun. 19, 2025

KBG6035700ab-Eaa
SA2506-0210

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company