

**CERTIFICATE OF ANALYSIS****LuciPac A3 Surface**

Product code :60361


Lot No.: 20250620Y

Expiry date: Sep. 19, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	13700	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	13900	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date Jul. 10, 2025KBG6036100ac-Eaa  
SA2507-0105Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company