

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484

Product name: **Lumitester Calibration Kit**

Serial Number: 25A031

DATA

for PD-30/PD-20

RLU value on the label:	1,140	RLU
(Std.: within 1100 \pm 400 RLU)		

RLU value obtained by a reference Lumitester PD-20: 1,140 RLU

Difference: 0.0 %

(Std.: within \pm 10 %)

for Smart

RLU value on the label:	1,430	RLU
-------------------------	-------	-----

RLU value obtained by a reference Lumitester Smart: 1,430 RLU

Difference: 0.0 %

(Std.: within \pm 10 %)

Date: Jul.10,2025

Signature: *Id. Kikuchi*
Hisako Kikuchi; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2507029