

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484
Product name: **Lumitester Calibration Kit**
Serial Number: 25A033

DATA

for PD-30/PD-20

RLU value on the label: 960 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 970 RLU
Difference: -1.0 %
(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,210 RLU

RLU value obtained by a reference Lumitester Smart: 1,230 RLU
Difference: -1.7 %
(Std.: within ± 10 %)

Date: Jul.10,2025

Signature: *Hisako Kikuchi*
Hisako Kikuchi; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2507031