

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484

Product name: **Lumitester Calibration Kit**

Serial Number: 25A034

DATA

for PD-30/PD-20

RLU value on the label: 1,710 RLU
(Std.: within 1100 \pm 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,720 RLU
Difference: -0.6 %
(Std.: within \pm 10 %)

for Smart

RLU value on the label: 2,320 RLU

RLU value obtained by a reference Lumitester Smart: 2,330 RLU
Difference: -0.4 %
(Std.: within \pm 10 %)

Date: Jul.10,2025

Signature: Id. Kikuchi
Hisako Kikuchi; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2507032