

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484

Product name: Lumitester Calibration Kit

Serial Number: 25A035

DATA

for PD-30/PD-20

RLU value on the label: 1,570 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,620 RLU

Difference: -3.2 %

(Std.: within ± 10 %)

for Smart

RLU value on the label: 2,020 RLU

RLU value obtained by a reference Lumitester Smart: 2,080 RLU

Difference: -3.0 %

(Std.: within ± 10 %)

Date: Jul.10,2025

Signature: Id. Kikuchi

Hisako Kikuchi; Manager

Quality Assurance Group, Edogawa

Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac

SA2507033