

CERTIFICATE OF ANALYSIS

Product code (Kikkoman): 60484

Product name: **Lumitester Calibration Kit**

Serial Number: 25A037

DATA

for PD-30/PD-20

RLU value on the label: 1,150 RLU
(Std.: within 1100 ± 400 RLU)

RLU value obtained by a reference Lumitester PD-20: 1,150 RLU

Difference: 0.0 %

(Std.: within ± 10 %)

for Smart

RLU value on the label: 1,470 RLU

RLU value obtained by a reference Lumitester Smart: 1,470 RLU

Difference: 0.0 %

(Std.: within ± 10 %)

Date: Jul.10,2025

Signature: Id. Kikuchi
Hisako Kikuchi; Manager

Quality Assurance Group, Edogawa
Quality Unit

Kikkoman Biochemifa Company

KBG60484-Eac
SA2507035