

# CERTIFICATE OF ANALYSIS

## Histamine Test

Product code :61341

Lot No.: 20250523A

Expiry date: Dec. 2027

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	100.4	100±5
Absorbance of reagent blank	0.012	≤0.04
Histamine dehydrogenase activity(U/vial)	2.21	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	98.5	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.



Hisako Kikuchi; Manager

Date Jul. 10, 2025

KBG6044101aa-Eaa  
SA2507-0108

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company