

CERTIFICATE OF ANALYSIS

LuciPac A3 Water

Product code :60365

Lot No.: 20250630W

Expiry date: Sep. 29, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100μL of 5pmol ADP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
When applied 100μL of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

J. d. Kikuchi
Hisako Kikuchi; Manager

Date Jul. 10, 2025

KBG6036500ac-Eaa
SA2507-0107

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company