

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361

Lot No.: 20250715Y

Expiry date: Oct. 14, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 $\mu$ L of 5pmol ADP Average Luminescence(RLU)	12300	$\geq 9100, \leq 16900$
When applied 100 $\mu$ L of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Aug. 01, 2025KBG6036100ac-Eaa  
SA2508-0020Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company