

CERTIFICATE OF ANALYSISLuciPac A3 Surface

Product code :60361


Lot No.: 20250725Y

Expiry date: Oct. 24, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12900	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	13100	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Aug. 25, 2025KBG6036100ac-Eaa
SA2508-0218Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company