

ELMEX LTD 

2-2, Ichigayasadohara-cho, Shinjuku-ku 162-0842 Tokyo, Japan
TEL: 81-3-5229-7911 FAX: 81-3-5261-4130
URL <http://www.elmex.co.jp>

Certificate of Analysis

Product	Sterilized Water	Code No.	MT-11PBS
Volume	9ml	Lot No.	5G1784
Manufactured	Apr 18, 2025	doc. No.	CAMT11PBS-20250516-7
Expiration	Mar. 2027	doc. Issued	May 16, 2025

Item	Specification	Result	Evaluation
Appearance: Liquid	clear, slight yellow liquid	clear, slight yellow liquid	passed
Appearance: Bottle	no visible particle	no visible particle	passed
Labels	correct position, correct printing	correct position, correct printing	passed
Net Weight	9.0±0.3g	9.0 g	passed
NaCl Content	0.85±0.1%	0.9%	passed
pH	7.0~7.4	7.3	passed
Leakage	no leak, -0.04Mpa hold 60min.	no leak	passed
Sterility Test	no bacteria (By TGC&PDA)	no bacteria	passed
Remarks			

ELMEX LTD

Authorised	Inspected
May 16, 2025 	May 16, 2025 