

**CERTIFICATE OF ANALYSIS****Histamine Test**

Product code :61341


Lot No.: 20250808A

Expiry date: Mar. 2028

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Quantifiability of histamine(%) (Ratio to liquid preparation)	99.8	100±5
Absorbance of reagent blank	0.011	≤0.04
Histamine dehydrogenase activity(U/vial)	2.20	≥ 0.64
Concentration of Histamine std. solution(%) (Ratio to the reference solution)	96.8	100±5

Storage condition : at 2-8°C protected from light

This is to certify that above information is a true and accurate copy of data obtained for this kit.

  
Hisako Kikuchi; Manager

Date Sep. 16, 2025KBG6044101aa-Eaa  
SA2509-0146Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company