

CERTIFICATE OF ANALYSISLuciPac A3 Water

Product code :60365


Lot No.: 20250905W

Expiry date: Dec. 04, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12500	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

Measured with Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Sep. 16, 2025KBG6036500ac-Eaa
SA2509-0153Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company