



CERTIFICATE OF ANALYSIS

LuciSwab ES 2.8-3.2


Product code :60356

Lot No.: 25YE2212B02

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Blank Luminescence Assay (RLU)		
Average Luminescence	passed test	≤ 20
Maximum Luminescence	passed test	≤ 25

Measured with LuciPac A3 Water and Lumitester Smart

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Sep. 16, 2025

KBG6035600ab-Eaa
SA2509-0152

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company