



CERTIFICATE OF ANALYSIS

LuciPac A3 Filter Assay

Product code :60376

Lot No.: 20250911A

Expiry date: Dec. 08, 2026

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
When applied 100 μ L of 5pmol ADP Average Luminescence(RLU)	12000	$\geq 9100, \leq 16900$
When applied 100 μ L of 5pmol AMP Average Luminescence(RLU)	12800	$\geq 9100, \leq 16900$

Mesured with Lumitester Smart.

This is to certify that above information is a true and accurate copy of data obtained for this kit.


Hisako Kikuchi; Manager

Date Sep. 26, 2025

KBG6037600ab-Eaa
SA2509-0264

Quality Assurance Group, Edogawa
Quality Unit
Kikkoman Biochemifa Company